



Molecular Human Genetics
JIWAJI UNIVERSITY GWALIOR

Gram: UNIVERSITY

Ref No. JU/17-18/MHG/1131

Date: 6-7-2017

SPEED POST

To,

Those vendors who are not registered but are authorized dealer of the equipments given below can also submit their quotation

Dear Sir,

Kindly send the quotations for following instruments latest by 24.07.2017 to the under signed by registered /speed post

SN	Item Description	Qty
1	Biosafety cabinet class 2, Type A-2 with HEPA filter for both main & exhaust, 30% air exhaust (Recirculatory) Model: MSW-164 Sterimax-4(A2) Size: 1200x600x600mm, SS 304 working table Exterior cabinet: MS duly powder coated Front sash – Toughened transparent glass, <i>magnetic gauge</i> Certifications required: CE, ISO & WHO-GMP certified Make: MAC / <i>Optics technology</i>	1 No.
2	REMI Refrigerated benchtop centrifuge, 1600RPM, Model: NEYA-16R LCD display of speed & time, imbalance detector, dynamic brake, safety lid lock, with voltage stabilizer and following accessories A. 4 Place swing out rotor, S4-175 B. 2 Place swing out rotor, S6-96MP C. Set of 4 inserts for 15ml falcon tube, I4-15F D. Set of 4 inserts for 50ml falcon tube, I1-50F E. Set of 4 inserts for 5/7ml vacutainer tubes, I7-5/7 F. Set of 4 metal buckets with biosafety cap for 175ml bottle, B175 Make: Remi	1 No.

Please attach following alongwith your bid:

- TIN no. / Sales Tax no. copy
- PAN No. copy
- EMD: Rs. 15000.00 to be submitted alongwith the offer

Zomh
Registrar
Jiwaji University
Gwalior (M.P.)

[Signature]
6/7/17

[Signature]
HEAD