



SOS IN BIOTECHNOLOGY, JIWAJI UNIVERSITY, GWALIOR

SPEED POST

Ref. No. JU/BT/18/138

Date : 28-09-2018

Coordinator
SOS in Biotechnology,
Jiwaji University, Gwalior- 474001

To,

Sub:- Quotation for Autoclave (vertical) and Serological water bath-reg.

Sir/Madam


Kindly send the quotations for following instrument latest by 19-10-2018 to the undersigned by Registered/speed post only.

S.N.	Item Description	Qty
01.	Autoclave (vertical) Inner and outer wall both made of stainless steel. Lid is made of thick machined stainless steel plate, Support ring and legs also made of stainless steel. Gasket is made from imported SILICONE rubber Volume : 40 liter Heater load : 2. Kw Inner Dimension : 300x500mm (DIA x HT) It must have PCO (pressure cut off) facility Accessory : One spare silicone gasket One spare heater	01
02.	Serological Water Bath Capacity : 18 liter and 04 reack Internal Dimension : 300x300x175mm Temperature : Ambient +5°C to 80°C ± 0.5°C Heat load : 1.0 ±1.0 KW Double walled with outer body made up of PCRC sheet duly powder coated, inner chamber and cover made up of Stainless Steel sheet (SS 304 grade). Thermostat controlled models. Accessories : Must provide 04 recks and one heating oil appropriate for this model	02

Please attach followings along with your bid;

- 01) Those venders who are not registered but are authorized dealer of the equipment can also submit their quotations.
- 02) EMD Rs. 5,000/- for Autoclave
- 03) Self certified copy of TIN no. /sales no.
- 03) Self certified copy of PAN no.
- 04) Price quoted should be F.O.R. destination i.e., Jiwaji University, Gwalior
- 05) Quote only on printed letter head format (no photocopy) & self certified.
- 06) Warranty: Two years from date of installation.


Registrar


Coordinator

IT cell
please load this
information on university
website.
