

SOS IN BIOCHEMISTRY, JIWAJI UNIVERSITY, GWALIOR

Ref. No. JU/BC/2017/ 110

Date: 25/05/2017

Head
SOS IN BIOCHEMISTRY,
JIWAJI UNIVERSITY, GWALIOR 474011

To,
M/S.
.....
.....
.....

Subject:- Quotation for pH Meter of SOS in biochemistry

You are requested to send sealed quotation (by Registered Letter/Speed Post Only) for the following items on or before **10/06/2017**

Specification for Items :-

S.no	Items	Brief Description
01.	pH Meter	Edge dedicated pH, 230: Resolution selectable from 0.01 & 0.0001 pH., Range: 2.00 to 16.000 pH., Accuracy: ±0.002 pH for 0.001 pH., Resolution: ±0.01 for 0.01 resolutions. Data Logging: Manual log-on demand, Manual log-on-stability, Interval Logging. Temperature readout (°C or °F), Automatic Temperature Compensation., CALL Check™ Indicators: Probe condition, Response time, Check buffer, Clean electrode. Sensor Check™ Indicators: Broken electrode, Closed junction. GLP data: Records. Date. Time, offset, slope and buffers used during calibration. Five-point calibration: A choice of seven preprogrammed buffers plus two selectable custom buffers. Calibration tag on screen: Identifies buffer Used for current calibration.

The Quoted price should be inclusive of the following :

- (04) All Taxes as per the Govt. rules
- (05) F.O.R. Department
- (06) Local transport and installation charges.

Important Note: - The Quotation reference should be clearly mentioned on top of the envelope.

Sincerely your

Head
SOS in Biochemistry
School of Studies in Biochemistry
Jiwaji University Gwalior

Atabulpariz

o/c
Atabulpariz
25/5/17