



**SCHOOL OF STUDIES IN BIOCHEMISTRY  
JIWAJI UNIVERSITY, GWALIOR 474 011**

Ref No. /

Gram : University  
Fax : 0751 2341450  
Date: 15.07.2017

To

Those vendors who are not registered but are authorized dealers of the equipment given below can also submit their quotation.

Dear Sir,

Kindly send the quotation for the following instrument latest by 31.07.2017 to the  
**Registrar, Jiwaji University, Gwalior 474 011**

**Kind Attention: Head, School of Studies in Biochemistry, Jiwaji University, Gwalior 474 011**  
by speed post/ registered post.

<u>S.No.</u>	<u>Item</u>	<u>Quantity</u>
01	Refrigerated Centrifuge Make : <b>Eppendorf</b> Model: <b>5430 R</b> Maximum Speed: 30,000g/ 17,500 rpm with a brushless motor Temperature range -10°C to 40°C [temp should be maintained to 4°C at maximum speed Memory To store 50 programs with 5 quick access program keys Acceleration time < 15s Braking time <15s Noise level < 54 db Flexibility To accommodate rotors for different formats of tubes(0.2ml PCR tube to spin column tubes , PCR and deep well plates Automation Automatic rotor recognition facility Program fast cooling with preset date and time Lid should have quick lock system for secure closing and opening User list Atleast 50 user of the quoted model Warranty Instrument should be CE certified have a IVD conformity Minimum 12 months from date of installation	01
02	Rotor 24 x b1.5/ 2.0ml High speed microcentrifuge tube [for use with centrifuge 5430R	01
	Rotor (quote as optional) 2 x microplates rotors for use with Centrifuge 5430R	01

Please attach the following along with your bid:

- i. Copy of TIN no./ Sales Tax No./ PAN no [if applicable]
- ii. EMD : Rs 15,000 to be submitted along with the offer

3/7/17  
[Registered]  
Registrar  
Jiwaji University  
Gwalior (M.P.)

*Neelam Bhandari*  
HEAD/Head  
School of Studies in Biochemistry  
Jiwaji University, Gwalior-474011 (M.P.)